



REQUEST FOR QUOTATIONS

No: RFQ-DGSTC 202106-01

For

DIGITAL SUBTRACTION ANGIOGRAPHIC SYSTEM

NAME OF COMPANY: STC (Dongguan) Company Limited (STC Dongguan, “Buyer”)

ADDRESS: 68 Fumin Nan Road, Dalang, Dongguan, Guangdong, China

Dear Sirs/ Madams,

Established in 2005 as STC’s nerve centre in Mainland China, STC Dongguan is an internationally recognized, world-class testing laboratory featuring state-of-the-art testing equipment and facilities of over 35,000 m². Offering extensive testing and certification services for a wide range of product categories, STC Dongguan is a Certification Body approved by the General Administration of Customs of the People’s Republic of China (GACC) and is the first non-governmental laboratory in China authorized to perform CCC testing and provide CCC certification services.

Since 2015, STC has established a preclinical medical device R&D test platform that meets international standards, and established a series of full-featured laboratories such as passive medical device laboratories and active medical device laboratories. To meet the pre-clinical R&D testing needs of medical device manufacturers, providing one-stop services, including chemical characterization, biocompatibility testing, large animal testing, microbiological testing, safety and performance verification, and electromagnetic compatibility testing. STC Medical Device Laboratories has been recognized by a wide range of domestic and international organizations, including IECEE, CE, FDA, NEMKO, INMETRO, CNAS, CMA, etc., and became the first Chinese OECD member of the GLP regulatory agency EMA (Entidad Mexicana de Acreditacion, ac) Review of certified medical device GLP laboratories.

We kindly request you to submit your quotation for DSA, short for Digital Subtraction Angiographic System.

When preparing your quotation, please be guided by the INSTRUCTION TO SUPPLIER attached hereto. Quotations shall be submitted in English and shall be marked with the note **“RFQ-DGSTC 202106-01 for DIGITAL SUBTRACTION ANGIOGRAPHIC SYSTEM”**

For interested Parties, please submit your Quotations to the below email address **on or before 30 Jun 2021**:

yenazhuang@stc.group

The Buyer is not bound to accept any quotation, nor award a contract, nor be responsible for any costs associated with a Supplier’s preparation and submission of a quotation, regardless of the outcome or the manner of conducting the selection process.

Thank you and we look forward to receiving your quotation.



Best regards,

STC (Dongguan) Company Limited

INSTRUCTION TO SUPPLIER

Customs clearance, if needed, shall be done by:	<input checked="" type="checkbox"/> Supplier
Address of Delivery Location	STC (Dongguan) Company Limited
Delivery Schedule	<input checked="" type="checkbox"/> Required
Preferred Currency of Quotation ₁	<input checked="" type="checkbox"/> United States Dollars, or <input checked="" type="checkbox"/> Chinese Yuan (CNY)
Value Added Tax on Price Quotation	<input checked="" type="checkbox"/> Must be inclusive of VAT and other applicable indirect taxes
After-sales services required	<input checked="" type="checkbox"/> Warranty on Parts and Labor for a minimum period of 1 year <input checked="" type="checkbox"/> Provision of Service Unit when pulled out for maintenance/ repair <input checked="" type="checkbox"/> Installation and adjustment of the goods <input checked="" type="checkbox"/> Instruction on Operation and Maintenance <input checked="" type="checkbox"/> Technical Support <input checked="" type="checkbox"/> Maximum 24 hours for urgent repair actions
Deadline for the Submission of Quotation	30 Jun 2021, 18:00 (Beijing time)
All documentations, including catalogs, instructions and operating manuals, shall be in this language	<input checked="" type="checkbox"/> English
Documents to be submitted	<input checked="" type="checkbox"/> Company profile (short info up to 1 page); <input checked="" type="checkbox"/> Detailed technical description of the offered goods: brochures, technical data sheet and specifications; <input checked="" type="checkbox"/> Certificate of origin (C / O) and a quality certificate (C / Q); <input checked="" type="checkbox"/> Statement or certificate of origin for the offered equipment; <input checked="" type="checkbox"/> Name and address of authorized service in China
Period of Validity of Quotes starting the Submission Deadline Date	<input checked="" type="checkbox"/> 12 months
Payment Term	<input checked="" type="checkbox"/> An advance payment is allowed up to a maximum of 30% of contract value.
Evaluation Criteria	<input checked="" type="checkbox"/> Technical responsiveness/Full compliance to requirements and lowest price. <input checked="" type="checkbox"/> Minimum 3 years of experience in the field; <input checked="" type="checkbox"/> Availability of certificates of quality and origin for the offered equipment; <input checked="" type="checkbox"/> Maximum delivery period not to exceed 06 months upon



	signature of contract; <input checked="" type="checkbox"/> Availability of authorized service in China and compliance with the after-sales services requirements.
The Buyer will award to:	<input checked="" type="checkbox"/> One and only one supplier
Type of Contract to be Signed	<input checked="" type="checkbox"/> Contract for Goods and/or Services
Conditions for Release of Payment	<input checked="" type="checkbox"/> Complete delivery and installation <input checked="" type="checkbox"/> Passing all Testing <input checked="" type="checkbox"/> Completion of Training on Operation and Maintenance
Annexes to this RFQ	<input checked="" type="checkbox"/> Technical Specifications of the Goods Required (Annex)
Contact Person for Inquiries	Yena Zhuang Email: yenzhuang@stc.group Contact number: +0769 8111 9888 # 8187



ANNEX: TECHNICAL SPECIFICATIONS

No.	Name of equipment/ Description	Code
	DIGITAL SUBTRACTION ANGIOGRAPHIC SYSTEM	
	Model:	
	Brand:	
	General information:	
	- Follow standards ISO 13485;	
	- Basic safety and performance standards: IEC 60601-1; IEC 60601-1-3; IEC 60601-2-28, IEC 60601-2-43, IEC 60601-1-2	
	- FDA 510(k) premarket notification;	
	- NMPA registration;	
	- Brand new 100%, can operate at once;	
1.1	INSTALLATION	
	<u>Floor-mounted stand</u>	
	<u>Be suitable to be installed in an operation room less than 6600mm*8000mm</u>	
	<u>Cover re-construction of the operation room and installation of the device</u>	
1.2	PRIMARY SYSTEM CONFIGURATION	
	<u>X-Ray Tube</u>	
	- Liquid metal bearing is preferred;	
	- 3.7MHU at least;	
	- Bronze-graphite;	
	- Efficient cooling system.	
	<u>Flat Detector</u>	
	- 16 bits Dynamic	
	<u>Patient Support</u>	
	- 270kg at least	
	<u>Generator</u>	
	<u>LCD Monitors in operating room</u>	
	<u>Exam light</u>	
1.3	ADVANCED FUNCTIONS	
	<u>Low dose technology</u>	
	<u>Navigation such as Roadmap, Overlay fade</u>	
	<u>Vascular analysis</u>	