

REQUEST FOR QUOTATIONS

No: RFQ-GDSTC 202412-01

For

TRANSESOPHAGEAL ECHOCARDIOGRAPHY

NAME OF COMPANY: STC (Guangdong) Company Limited (STC Guangdong, “Buyer”)

ADDRESS: No.68, Fumin South Road, Dalang, Guangdong, Guangdong, China

Dear Sirs/ Madams,

Established in 2005 as STC’s nerve centre in Mainland China, STC Guangdong is an internationally recognized, world-class testing laboratory featuring state-of-the-art testing equipment and facilities of over 35,000 m². Offering extensive testing and certification services for a wide range of product categories, STC Guangdong is a Certification Body approved by the General Administration of Customs of the People’s Republic of China (GACC) and is the first non-governmental laboratory in China authorized to perform CCC testing and provide CCC certification services.

Since 2015, STC has established a preclinical medical device R&D test platform that meets international standards, and established a series of full-featured laboratories such as passive medical device laboratories and active medical device laboratories. To meet the pre-clinical R&D testing needs of medical device manufacturers, providing one-stop services, including chemical characterization, biocompatibility testing, large animal testing, microbiological testing, safety and performance verification, and electromagnetic compatibility testing. STC Medical Device Laboratories has been recognized by a wide range of domestic and international organizations, including IECEE, CE, FDA, NEMKO, INMETRO, CNAS, CMA, etc., and became the first Chinese OECD member of the GLP regulatory agency EMA (Entidad Mexicana de Acreditacion, ac) Review of certified medical device GLP laboratories.

We kindly request you to submit your quotation for TEE, short for Transesophageal Echocardiography

When preparing your quotation, please be guided by the INSTRUCTION TO SUPPLIER attached hereto. Quotations shall be submitted in English and shall be marked with the note **“RFQ-GDSTC 202412-01 for Transesophageal Echocardiography”**

For interested Parties, please submit your Quotations to the below email address on or before 20 Jan 2025:

eugenehuang@stc.group

The Buyer is not bound to accept any quotation, nor award a contract, nor be responsible for any costs associated with a Supplier’s preparation and submission of a quotation, regardless of the outcome or the manner of conducting the selection process.

Thank you and we look forward to receiving your quotation.

Best regards,

STC (Guangdong) Company Limited

INSTRUCTION TO SUPPLIER

| | |
|---|--|
| Customs clearance, if needed, shall be done by: | <input checked="" type="checkbox"/> Supplier |
| Address of Delivery Location | STC (Guangdong) Company Limited |
| Delivery Schedule | <input checked="" type="checkbox"/> Required |
| Preferred Currency of Quotation | <input checked="" type="checkbox"/> United States Dollars, or <input checked="" type="checkbox"/> Chinese Yuan (CNY) |
| Value Added Tax on Price Quotation | <input checked="" type="checkbox"/> Must be inclusive of VAT and other applicable indirect taxes |
| After-sales services required | <input checked="" type="checkbox"/> Warranty on Parts and Labor for a minimum period of 1 year <input checked="" type="checkbox"/> Provision of Service Unit when pulled out for maintenance/ repair <input checked="" type="checkbox"/> Installation and adjustment of the goods <input checked="" type="checkbox"/> Instruction on Operation and Maintenance <input checked="" type="checkbox"/> Technical Support <input checked="" type="checkbox"/> Maximum 24 hours for urgent repair actions |
| Deadline for the Submission of Quotation | 20 Jan 2025, 17:30 (Beijing time) |
| All documentations, including catalogs, instructions and operating manuals, shall be in this language | <input checked="" type="checkbox"/> English |
| Documents to be submitted | <input checked="" type="checkbox"/> Company profile (short info up to 1 page); <input checked="" type="checkbox"/> Detailed technical description of the offered goods: brochures, technical data sheet and specifications; <input checked="" type="checkbox"/> Certificate of origin (C / O) and a quality certificate (C / Q); <input checked="" type="checkbox"/> Statement or certificate of origin for the offered equipment; <input checked="" type="checkbox"/> Name and address of authorized service in China |
| Period of Validity of Quotes starting the Submission Deadline Date | <input checked="" type="checkbox"/> 12 months |
| Payment Term | <input checked="" type="checkbox"/> An advance payment is allowed up to a maximum of 30% of contract value. |
| Evaluation Criteria | <input checked="" type="checkbox"/> Technical responsiveness/Full compliance to requirements and lowest price. <input checked="" type="checkbox"/> Minimum 3 years of experience in the field; <input checked="" type="checkbox"/> Availability of certificates of quality and origin for the offered equipment; <input checked="" type="checkbox"/> Maximum delivery period not to exceed 3 months upon signature of contract; <input checked="" type="checkbox"/> Availability of authorized service in China and compliance with the after-sales services requirements. |
| The Buyer will award to: | <input checked="" type="checkbox"/> One and only one supplier |
| Type of Contract to be Signed | <input checked="" type="checkbox"/> Contract for Goods and/or Services |
| Conditions for Release of Payment | <input checked="" type="checkbox"/> Complete delivery and installation <input checked="" type="checkbox"/> Passing all Testing <input checked="" type="checkbox"/> Completion of Training on Operation and Maintenance |

| | |
|------------------------------|--|
| Annexes to this RFQ | <input checked="" type="checkbox"/> Technical Specifications of the Goods Required (Annex) |
| Contact Person for Inquiries | Eugene Huang Email: eugenehuang@stc.group Contact number: +0769 8111 9888 |

ANNEX: TECHNICAL SPECIFICATIONS

| No. | Name of equipment/ Description | Code |
|------------|---|------|
| | Transesophageal Echocardiography | |
| | Model: | |
| | Brand: | |
| | General information: | |
| | - Follow standards ISO 13485; | |
| | - FDA premarket notification; | |
| | - NMPA registration; | |
| | - Brand new 100%, can operate at once; | |
| | - Import the whole machine in original package. | |
| 1.1 | PRIMARY SYSTEM CONFIGURATION | |
| | <u>Architecture</u> | |
| | - The host supports the pure wave matrix transesophageal technology; | |
| | - Pure wave matrix probe frequency: 2-7 MHz; | |
| | - Real-time 3D imaging technology; | |
| | - Real-time biplane imaging technology is supported; | |
| | - Supports 2D ICE probes. | |
| | <u>Interface</u> | |
| | - 15.0 inch high resolution display with wide viewing angle; | |
| | - ≥ 500 GB hard drive; | |
| | - Internal DVD RW drive. | |
| | <u>Imaging Mode</u> | |
| | - Live 3D TEE and Live 3D TEE color flow; | |
| | - 2D、M-mode、Anatomical M-mode、Color M-mode; | |
| | - Color Power Angio (CPA) ; | |
| | - Pulsed Wave Doppler、Continuous Wave Doppler; | |
| | - Invert and Color Invert、Color compare mode、Dual mode、Duplex for simultaneous 2D and Doppler; | |
| | - Live Compare、Tissue Harmonic Imaging (THI)、Reconstructed zoom with pan (read zoom)、Write zoom; | |
| | - Pulse Inversion Harmonic imaging、Adaptive Doppler、Adaptive Color Doppler、Color Tissue Doppler imaging、Pulsed Wave Tissue Doppler imaging; | |
| | - Active Native Data – manipulation of image data. | |
| | <u>DICOM Networking</u> | |
| | - Ethernet @100Mb/second Includes DICOM wired and wireless "G" and "N"; | |
| | - Provides DICOM 3.0 network print and store and storage commitment. | |
| 1.3 | <u>Transducer OPTIONS</u> | |
| | - Transesophageal Transduce; | |
| | - Transthoracic Transduce; | |
| | - Transabdominal Transduce. | |